

## EC Type-Examination



No. CE 561481

Issued to:

**3M United Kingdom Plc**  
**3M Centre**  
**Bracknell**  
**Berkshire**  
**RG12 8HT**  
**United Kingdom**

In respect of:

**Respiratory protective devices to EN 149:2001+A1:2009:-  
Filtering half masks to protect against particles.  
Model: 3M 9926 Particulate Respirator.  
See pages 2 and 3 for details.**

On the basis of our examination, under the requirements of Council Directive 89/686/EEC "Personal Protective Equipment" Article 10, EC Type-Examination.

For and on behalf of the British Standards Institution, a Notified Body for the above Directive (Notified Body Number 0086):

A handwritten signature in black ink, appearing to read 'David Ford', written over a horizontal line.

David Ford, Director, Healthcare and Testing Services

First Issued: 11 May 2010

Date: 11 May 2010

Page: 1 of 3

## EC Type-Examination

No. CE 561481

3M United Kingdom Plc

Bracknell

### Product Details

Product name:	Particulate Respirator.
Models:	3M 9926 (Standard Industrial Product).
Product type:	Filtering half mask to protect against particles.
Technical Specification:	Harmonized European Standard EN 149:2001+A1:2009.
EN 149 classification:	FFP2. Single shift use - symbol NR. Dolomite clogging option - symbol D.

The product assessments were based on BS EN 149:2001+A1:2009, the English language version of EN 149:2001+A1:2009, respiratory protective devices – filtering half masks to protect against particles, both documents incorporating Corrigendum dated July 2002.

The valved respirator is designed to protect against solid and non-volatile liquid particles, is held onto the face by a pair of elasticated straps and is single shift device.

### Packaged variants

In addition to the product referenced on this Certificate, the standard industrial product may also be sold as market specific packaged variants. Such variants will be identified by an alpha suffix added to the model reference and the Technical File will be updated with the appropriate information.

### Kits and packouts

The product referenced on this Certificate may also be combined with other 3M products into a kit or packout. There will be no change to the product but the user information may vary. In such instances the Technical file will be updated with the appropriate information.

First Issued: 11 May 2010

Date: 11 May 2010

Page: 2 of 3

## EC Type-Examination

No. CE 561481

3M United Kingdom Plc

Bracknell

### Certification Administration Details

Technical File Reference: TF0166

### Certificate Amendment Record and internal BSI report related to this Certificate:

Issue date	Comments	BSI Report No.
May 2010	First issue of Certificate for the 3M 9926 assessed to EN 149:2001+A1:2009 (1).	0086:7465691/8

(1) Prior to the issue of this Certificate the product was Certified to EN 149:2001, Certificate CE 58494 refers.

### Certificate validity

The Certificate holder is responsible for ensuring that the Notified Body is advised of changes to any aspect of the overall processes utilised in the manufacture of the product, failure to do so could invalidate the Certificate in respect of product manufactured following the introduction of such changes.

The validity of the Certificate is also dependent on the maintenance of the EC quality of production by monitoring system, PPE Article 11B, as referenced on BSI Certificate 00347.

First Issued: 11 May 2010

Date: 11 May 2010

Page: 3 of 3

raising standards worldwide™

